### DIAMETER 5" NOMINAL

### 5EG2P


# Oscilloscope Tube

ELECTROSTATIC FOCUS. ELECTROSTATIC DEFLECTION.

## DATA

DATA	
GENERAL:	
Heater: Voltage 6.3	. a.c. or d.c. volts.
Current 0.6	. amp.
Direct Inter-electrode Capacitances:	
Modulator to all other electrodes .	. 12.0μμf.
Each X Plate to all other electrodes .	$5.0\mu\mu f$ .
Each Y Plate to all other electrodes .	. 5.0µµf.
Deflector Plates XI to X2	2.3µµf.
Deflector Plates Y1 to Y2	. 2.3µµf.
Screen:	( (
Fluorescence	. Green.
Persistence	. Short
(10 m. sec. min./100m. sec. max. for 1	
Focussing Method	. Electrostatic.
Deflecting Method	. Electrostatic.
Deflecting Method	
Overall Length	. 425 ± 9 mm. . 136.5 mm.
Greatest Diameter of Bulb  Minimum Useful Screen Diameter	. 130.3 mm.
Minimum Useful Screen Diameter	. 114 mm.
Mounting Position	<ul><li>Any.</li><li>English BSS448/CT2.</li></ul>
Anode Cap	. English BSS448/C12.
Base	. 11 Pin Magnal.
Pin 1—Heater. Pin 2—No connection. Pin 3—No connection. Pin 4—Anode 2. Pin 5—No connection. Pin 6—No connection.	Pin 7—Anode 1 and Anode 3. Pin 8—No connection. Pin 9—No connection. Pin 10—Modulator. Pin 11—Heater and Cathode. Cap—Anode 4 P.D.A. BSS448/CT1 caps on neck connect to X1, X2, Y1 and Y2.
Typical Operating Conditions: Anode 1 and 3 (2500 volts max.).	. 2000 volts.
Anode 2	450/570 volts.
Anode 4 (5000 volts max.)	. 4000 volts.
Modulator volts for cut-off	105 volts max.
Wiodulator voits for cut-on	105 voits max.
75 G 11 G 11 11	
Deflection Sensitivity:	mm /volt
•	mm./volt.
X Plate	mm./volt 0.22 to 0.33 0.22 to 0.33.





#### ALL SIZES IN MILLIMETRES.

- Note 1. The angle between the trace produced by Y1, Y2 and a plane through the tube axis, spigot key and the P.D.A. Cap, may vary by an angular tolerance of 10°. The spigot key is on the same side of the tubes as the P.D.A. Cap.
- Note 2. The angle between the trace produced by X1 and X2 and the trace produced by Y1 and Y2 is 90° ± 3°.
- Note 3. The undeflected focused spot will fall within a circle having a 7 m.m. radius concentric with the centre of the tube face.
- Note 4. When viewing the screen with the tube positioned such that the spigot key is uppermost, a positive voltage applied to the terminal X1 will deflect the spot to the left and a positive voltage applied to the terminal Y1 will deflect the spot upwards.